Erba Lachema s.r.o.

ADVISORY NOTICE

Product name: CREATININE ENZYMATIC

Pack name: CREA ENZ 200

Cat. no. XSYS0085

Dear Customer

We hereby inform you that the improved kit for determination of CREATININE ENZYMATIC has been launching.
The following changes have been implemented in the kit:

1. Change of composition

The principle of the determination is the same. The content of active ingredients is slightly different. The working instruction was changed. Please, read the working instruction carefully.

2. Assay procedure

The assay procedure (volume of reagents, ration reagents/samples, wavelenght,…) is without any change. There measuring range (technical limits) is changed (LOD and linearity). Please, upgrade the application in your analyser (new SW) – you can contact your local service team to help you.

In case you have any quastions, please do not hesitate to contact us.

Best regards,

Erba Lachema s.r.o.
Technical Product Management

Brno 26.03.2021

Erba Lachema s.r.o.
Karásek 2219/1d, 621 00 Brno
DIČ: CZ26918846 -22-