Erba Lachema s.r.o.

ADVISORY NOTICE

Product name: Erba Control N Plus
Erba Control P Plus

Cat. no.: EHL00016
EHL00017

LOT: 2010012
2010013

Dear customer,

We hereby inform you that we have extended the value sheet with missing values.

Erba Control N Plus LOT 2010012	Unit	ECL 105 / 412		ECL 760	
		Value	Range	Value	Range
Factor IX Deficient Plasma	%	150.3	120.2–180.3	156.5	125.2–187.8
Protein C	%	117.4	93.9–140.9	122.1	97.7–146.5

Erba Control P Plus LOT 2010013	Unit	ECL 105 / 412		ECL 760	
		Value	Range	Value	Range
Factor IX Deficient Plasma	%	67.0	53.6–80.4	68.7	55.0–82.5
Protein C	%	62.9	50.4–75.5	75.7	60.6–90.8

Best regards,

Erba Lachema s.r.o.
Technical product management

Erba Lachema s.r.o.
Karásek 1d, 621 33 Brno
DIČ: CZ26918846 -2-

Brno 26th August 2021