ADVISORY NOTICE

XL TEST LOADER/OPEN SYSTEM CARDS registration failure – extended notice, version 2

Dear customer,

we would like to inform you that we have identified functional issue with test loader and open system cards for XL analyzers. Affected products **would fail to register once used on XL analyzers**.

These cards are used to open closed system of XL chemical analyzers used for automatic testing of routine clinical trials. Without functional cards the system cannot be opened.

Bellow, please, find the list of affected products and lots:

Product	REF	Lot No.
XL 200 TEST LOADER CARD	10020737	2102062
XL CC systems TEST LOADER CARD	10020739	2102196
XL CC systems TEST LOADER CARD	10020739	2103031
XL 1000 OPEN SYSTEM CARD	10020769	2103061
XL 640 OPEN SYSTEM CARD	10020770	2103062
XL 200 OPEN SYSTEM CARD	10020771	2108160
XL 180 OPEN SYSTEM CARD	10020772	2103064

We kindly ask you to secure provided cards, discard it and provide us confirmation of the action taken.

Please accept our sincere apologies for the inconvenience caused. In case you have any questions, please do not hesitate to contact us.

Best regards,

Erba Lachema s.r.o.
Karásek 1d, 621 33 Brno
DIČ: CZ26918846

Ing. Radim Herrmann
Quality Manager
Erba Lachema s.r.o.

Brno, 8th December 2021

Erba Lachema s.r.o., Karásek 2219/1d, 621 00 Brno, Czech Republic
Identification number: 269 18 846. Tax identification number: CZ26918846
Incorporated in the Commercial Register maintained by the Regional Court in Brno, Section C, insert 45458
Tel.: +420 517 077 111, e-mail: diagnostics@erbamannheim.com, www.erbalachema.com