Erba Lachema s.r.o.

ADVISORY NOTICE


12_2023_LAURA XL Rack stickers and barcodes

Dear customer,

We would like to inform you that we have identified omission on Erba Lachema side linked with accessory box of following LAURA XL instruments you have received. Below mentioned instruments are missing barcode stickers on tube racks in accessory box therefore racks wouldn't be identified by your analyser within usage.

We would like to apologize for any inconvenience cause by this mistake and we would like to ask you to stick labels on racks by yourself. These labels will be sent to you free of charge.

Please attach the barcodes and rack no. stickers according to following scheme. Please pay attention on rack numbers to apply stickers accordingly:



Applied for:

INSTRUMENT: LAURA XL cat.no.INS00065

Serial numbers:
0381-X2-211022
0382-X2-211022
0385-X2-211022
0386-X2-071122
0387-X2-071122
0388-X2-071122
0390-X2-071122

In case you have any additional question, please do not hesitate to contact us.

Best regards,

Erba Lachema s.r.o.
Technical product management

Erba Lachema s.r.o.
Karásek 2219/1d, 621 00 Brno
DIČ: CZ26918846 -7-

Brno 12th May 2023 Ref.No.AN_12_2023