Erba Lachema s.r.o.

ADVISORY NOTICE

14_2021_EC 90_SW version V1.00.1

Dear Customers,

We hereby inform you, that the device software of the EC 90 Analyzer must be updated to SW version V1.00.1. issued 16th July 2021. The SW version V1.00.1 is already included from the device SN 3154-2-202103 onwards.

For SW version V1.00.1. update, please contact your local service engineer or distributor to set up new SW version from web side ERBA Academy.

Applied for:

INSTRUMENT: EC 90 Analyzer cat. no. INS00079

In case you have any additional questions, please do not hesitate to contact us.

Best regards,

Erba Lachema s.r.o.
Technical product management

Brno 27th July 2021

Ref.No.: AN_2021-14

Erba Lachema s.r.o.
Karásek 1d, 621 33 Brno
DIČ: CZ26918846 -2-