Erba Lachema s.r.o.

ADVISORY NOTICE

25_2024_XL-200 Plus_SW_v2022.04

Dear customer,

We hereby inform you that the XL-200 Plus device software has been updated to SW version 2022.04 issued on 24th April 2024.

XL-200 Plus SW_v2022.04 implements bug fix and new feature:

- Issue resolved: ISE> Run gets stopped due to 'recovering R1 Skipped and reschedule' 'RCT tray rotational error' and CRU interlock error during run.
- SW_v2022.04 supports Multiple Screen Resolution, that are higher then1152 x 864. The best-viewed resolution is 1152*864 (recommended).

SW version 2022.04 is already included in devices produced from SN 20P0221 onwards.

For SW version 2022.04, please contact your local service engineer or distributor to set up the new SW version from the web site ERBA Knowledge Base.

Applied for:

INSTRUMENT: XL-200 Plus cat.no: INS00001, INS000002

In case you have any additional questions, please do not hesitate to contact us.

Best regards,

Erba Lachema s.r.o.

Erba Lachema s.r.o.
Karásek 2219/1d, 621 00 Brno
DIČ: CZ26918846 -7-

Brno, 19th August 2024 Ref. No.AN_25_2024