Erba Lachema s.r.o.

ADVISORY NOTICE

05_2026_XL MULTICAL re-calibration

Cat. No.:	Pack Name	Packaging (Content)
XSYS0034	XL MULTICAL 4x3	R1: 4 × 3 mL
XSYS0122	XL MULTICAL 10x3	R1: 10 × 3 mL

Dear customers,

With a mission to continuously enhance user experience and satisfaction, we would like to inform you about the re-calibration and traceability changes of calibrator XL MULTICAL.

As part of the IVDR validations, analytes in the calibrator were recalibrated using the latest reference methods / materials.

Significant differences in values were found only for ALT, amylase and uric acid. Advisory notices for these analytes with updated values can be found on Erba website:

2025_57_Re-assignment of alanine aminotransferase (ALT/GPT)
2026_03_ Re-assignment of amylase (AMY)
2026_04_ Re-assignment of uric acid (UA)

As a result of this process, a design modification to the value sheet for XL MULTICAL was also made. Several analytes that previously had united values (amylase, glucose, uric acid) have now been divided into separate rows for each determination method.
At the same time, the abbreviations for individual analytes have been modified to match the approved applications in our XL analyzers. This process will make the values in the sheet clearer and minimize the risk of error when manually entering values into the analyzers.

Based on these changes, the certificate of traceability was also updated. The latest version of the new lot of XL MULTICAL is available from your local representatives upon request.

In case you have any additional questions, please do not hesitate to contact us.

Best regards,

Erba Lachema s.r.o.
Technical Product Management

Brno, 21st January 2026

Erba Lachema s.r.o.
Karásek 2219/1d, 621 00 Brno
DIČ: CZ26918846 -7-

Ref. No.AN_05_2026